TITLE OF THE INVENTION

Optochemical Sensing with Multi-Band Fluorescence Enhanced by Surface Plasmon

Resonance.

Inventors: **Malak, Henryk** (Ellicott City, MD) and **Bogdanov, Valery** (Lawrenceville, NJ)
Assignee: American Environmental Systems, Inc.

Provisional Application No.: 60/446,096

Original Filing Date: 02/10/2003

Non-Provisional Filing Date: 09/04/2003

Current U.S. Class: **435/6**; 435/174; 435/287.2; 435/7.9; 436/525; 436/94

U.S. Class at Publication: **435/6**; 435/7.9; 436/525; 436/94; 435/287.2; 435/174

Intern'l Class: C12Q 001/68; G01N 033/53; G01N 033/542; G01N 033/553

CROSS REFERENCE TO RELATED APPLICATIONS

This application is related to U.S. Provisional Patent Application Number 60/446,096,

entitled "Optochemical Sensing with Multi-Band Fluorescence Enhanced by Surface

Plasmon Resonance" filed February 10, 2003, which is herein incorporated by reference.

STATEMENT REGARDING FEDERALLY SPONSORED RESEARCH

There is NO claim for federal support in research or development of this product.

REFERENCES CITED

The following are patents found that may be associated with this information.

U.S. Patent Documents

US 5,866,433   February 2,1999   Schalkhammer, et al.

US RE37,412   October 16, 2001   Schalkhammer, et al.

BACKGROUND OF THE INVENTION

This invention relates to an optochemical sensing of materials for molecular identification and measuring the concentration of one or more analytes in the sample. Optochemical sensing is based on reading an optical signal generated by a sensor interacting with an analyte. Fluorescence sensors as a group are the most sensitive optochemical sensors that utilize a fluorescence signature of sensor and/or analyte (excitation and/or emission spectra, intensity, lifetime, polarization) to identify materials with high specificity. The invention discovers new principles of fluorescence sensing that improves fluorescence sensor specificity, sensitivity and response time, dramatically. The invention will be applied to life sciences, biomedicine, defense, and material science research and application.

DESCRIPTION OF THE PRIOR ART

Schalkhammer, et al., US 5,866,433. An optochemical fluorescence sensor with a biorecognitive layer for measuring the concentration of one or more analytes in a sample is provided with at least one island layer that is applied on a sensor substrate. The islands of the island layer are in the form of electrically-conductive material and have a diameter

of less than 300 nm, the biorecognitive layer being directly applied on the island layer or

bound via a spacer film. In addition, an analyte-specific fluorescent compound is

provided which may be added to the sample or is provided in the sensor itself. The

biorecognitive layer can bind the analyte to be measured directly or by means of analyte-

binding molecules, the originally low quantum yield of the fluorescent compound

increasing strongly in the vicinity of the island layer.

Schalkhammer, et al., US RE37,412. An optochemical sensor for measuring

concentrations of analytes is provided with a reactive matrix preferably made of

polymeric material capable of swelling. Further provided are a mirror layer and a layer of

a plurality of discrete islands that are electrically conductive, between which layers the

reactive matrix is positioned, the diameter of the islands being smaller than the

wavelength of the light employed for monitoring and evaluation.

SUMMARY OF THE INVENTION

The objective of the invention is to provide a method for the optochemical fluorescence

sensing of fluorophores and/or analytes, which will allow very sensitive, up to single

molecule specific, fast identification of the fluorophore and/or will aid in measure of

analyte concentration.

In the invention this objective is achieved by measurement and analysis of multiband

emission of the fluorophore enhanced by electro-magnetic fields of surface plasmon

resonance (SPR) from nearby metal nanoparticles.

Multiband fluorescence includes emission bands from low excited state (LES) and higher

excited states (HES) of the same analyte molecule (Fig.1). As a rule, HES emission is

difficult to measure due to ultra-fast HES nonradiative decay and very low quantum yield

of HES fluorescence (< 1%, typically). All current fluorescence sensors employ

measurement at a single-band LES fluorescence of analyte.

The invention expands an analytical capacity of conventional, single-band fluorescence

spectroscopy and sensing through implementation of the method to measure enhanced

multi-band - HES and LES fluorescence bands of the same fluorophore. The method

provides a band-selective enhancement of a low quantum yield emission of HES

fluorescence that leads to easy-to-detect multi-band fluorescence sensing.

The invention employs a dependence effect of fluorophore emission rate enhancement by

nearby metal (silver, gold) nanoparticles on quantum yield of fluorophore (Q). If the

fluorophore is positioned inside SPR evanescent wave zone, strong electromagnetic fields

generated by surface plasmons of nanoparticles, enhance fluorophore absorption and

emission rate 100s-1000s folds, respectively. This leads to an enhanced fluorescence

quantum yield and measured intensity. However, enhanced quantum yield is limited by a

maximum Q value equal to 100%. This limitation results in a relatively low efficiency of

intensity enhancement for a fluorophore with high Q value; but, it does not place a

practical limit on the enhanced emission intensity of a fluorophore with low Q value.

Experimental data confirms this conclusion: the emission intensity measured for a series

of fluorophores in the vicinity of metal nanoparticles was greatly increasing for the

decreasing values of Q for fluorescence (Fig.2, graph based on data published by J. R.

Lakowicz et al. "Radiative Decay Engineering. 2. Effects of Silver Island Films on

Fluorescence Intensity, Lifetimes, and Resonance Energy Transfer", Anal. Biochem.,

301:261 (2002)). Therefore, SPR-mediated growth of emission intensity is expected to be

high for excited electronic state with low quantum yield (HES) and low for state with a

high quantum yield (LES) of the same fluorophore. As a result, the fluorescence

intensities from the HES and LES would reach comparable levels. Thus, at enhanced

fluorescence with metal nanoparticle, HES fluorescence could be used as an additional,

measurable optical signature of fluorophore.

The effect of fluorescence enhancement depends on a distance between fluorophore and

nanoparticle. As the distance increases, enhance fluorescence signal decreases, and a

significant enhancement is available for a fluorophore positioned inside an evanescent

wave zone of surface plasmon, only. On the other side, direct metal-fluorophore contact

energy transfer quenches fluorescence completely. Ultra-thin dielectric barrier-spacer

(plastic, polymer or SiO_2 layer with thickness above 10 nm) between nanoparticle and

fluorophore eliminates quenching and causes significant fluorescence enhancement. The

upper limit of the layer thickness for fluorescence enhancement depends on the depth of

evanescent light penetration in the optochemical sensor. According a recent paper of

Horváth and et al., a refractive index of planar waveguide can increase the penetration of

evanescent zone up to a few microns (Horváth et al. "Optical waveguide sensor for on-

line monitoring of bacteria", Optics Letter, 28, 1233 (2003)).

Additionally, molecular probes designed to capture analyte in addressable location

(microarray) could be developed on spacer surface. It leads to a sensor design capable of

simultaneous and highly parallel multi-band fluorescence sensing of analytes.

Thus, a sensor in the invention comprises of:

a) a support with designed refractive index of used material

b) a thin film of metal nanoparticles or metal islands placed on a surface of support

c) a dielectric or biorecognitive barrier with thickness more than 10 nm placed on a

nanoparticle coated substrate

d) analyte probes attached or not attached to an external barrier surface (microarray or

other format)

e) an excitation source: optical illumination or chemical reaction

The invention can also be applied to sensor design based on single metal nanoparticle or

an assembly of metal nanoparticles, where each nanoparticle is coated or not coated with

dielectric or biorecognitive barrier, dispersed in medium, cells, or other sensing materials

and excited with light causing surface plasmon enhanced emission of surrounding

fluorophores.

Hyperspectral imaging is a preferable method to measure and analyze fluorescence of

analytes immobilized on a sensor surface. Fig. 3 shows a possible schematic of proposed

sensor and Fig. 4 shows a hyperspectral imager (HIS) employing the effect of slit-free,

optical-rotation dispersion on polychromatic radiation (P. Herman et al. "Compact

hyperspectral imager for low light applications" *SPIE Proc.* 2001, 4259, pp. 8-16).

However, other optical techniques can be also applied with optochemical multiband

enhanced fluorescence sensing, like time-resolved spectroscopy, fluorescence

polarization, fluorescence recovering after photobleaching, fluorescence resonance

energy transfer surface, enhanced multiband Raman scattering (but not limited to them).

The multi-band enhanced emission can be generated by electromagnetic radiation source

in single, and multi- photon and/or nonlinear optical modes of excitation. It can be also

generated by chemiluminescence, electro-optically, electrochemically and other

luminescence techniques. In all of these methods, band-selective intensity enhancement

leads to comparable intensity HES and LES bands.

Thus, hyperspectral detection and other above mentioned techniques could be used for

optochemical sensing employed multi-band enhanced emission.

BRIEF DESCRIPTION OF THE DRAWINGS

Fig.1. Schematic diagram of the fluorophore electronic states, processes (left) and

fluorescence spectra (right). 1 and 2 – one photon absorption/LES and HES population,

1+3- two-photon step wise absorption/HES population, 4- LES fluorescence, 5- HES

fluorescence, 6 – LES nonradiative decay, 7- HES nonradiative decay.

Fig. 2. Dependence of enhanced fluorescence intensity with nearby silver nanoparticle on fluorophore quantum yield.

Fig. 3. Schematic diagram of the proposed sensor and hyperspectral optical setup.

SENSOR (insert). Metal nanoparticles (tens of nanometers in diameter) are placed on the surface of glass substrate (prism). Nanoparticle layer is coated with a 10-100s nm thick dielectric layer (polymer or SiO_2) to create a physical barrier between the metal particles and a fluorophore. Microarray of analyte captured spots is attached to a surface of dielectric layer. The excitation can be delivered *via* evanescent wave coupling using the effect of total internal reflection at the prism surface.

OPTICAL SET-UP. The entire microarray can be illuminated with laser pulses at two different wavelengths. To produce both LES and HES signatures, the sample is simultaneously illuminated by two nanosecond laser pulses at different wavelengths, for example 4[th] harmonics (266 nm) and fundamental (1064 nm) wavelength of Nd:YAG laser. The conventional (LES) fluorescence spectrum will be acquired following single-photon excitation at 266 nm. To obtain HES fluorescence spectrum, two-photon resonant (step-wise) excitation is used. In the first step, LES is populated when the molecules in their ground electronic state absorb a photon at 266 nm. In the second step the excited molecules in LES absorb the second photon at 1064 nm; this results in population of HES. The measured HES fluorescence spectrum is blue-shifted compared to the LES fluorescence. To obtain the full analyte optical signature (LES + HES fluorescence),

many other combinations can be used in the step-wise excitation. A Nd:YAG laser

equipped with a standard set of nonlinear crystals can generate pulses at the fundamental

frequency plus four harmonics.

In this example, the output a Q-switched Nd:YAG laser (5 ns pulses, up to 100 Hz

repetition rate), consists of the fundamental (1064 nm) and 2^{nd} harmonics (532 nm)

and/or 3^{rd} harmonics (355 nm), and/or 4^{th} harmonics (266 nm). This multitude of

wavelength provides a high degree of flexibility in detection of practically any

organic/inorganic matter. The fundamental output is divided into two beams by means of

an 60/40 beam splitter. The 40% fraction of the 1064 nm beam passes through a an

assembly of nonlinear crystals and is converted into the harmonics which are directed

into the total internal reflection (TIR) prism made of fused silica. The harmonics

illuminate the glass-sensor interface at the critical angle and excite the bio-agent

fluorophores attached (captured) to the microarray *via* evanescent wave illumination. The

remaining 60% of the fundamental (1064 nm) enters the TIR prism from the opposite

prism side and overlaps with the harmonics beam at the glass-sensor interface. A shutter

placed in the fundamental beam controls the excitation scheme by blocking passing

the1064 nm radiation. Microarray emission is collected by infinity-corrected lens and

transmitted through laser cut-off filter and the HSI module. An imaging lens produces a

microarray spectral image which is then captured by a cooled CCD array.

Fig. 4. Schematic diagram of hyperspectral imager (HIS) module in Fig. 3.

This module utilizes the effect of optical rotation dispersion (ORD) on polychromatic

light. It consists of a polarization rotator – an optically active medium (crystalline

quartz) placed between a pair of polarizers with their transmission axes aligned parallel to each other. The polarization direction of linearly polarized input light rotates during propagation through the rotator and the rotation angle depends on the wavelength and the rotation power of the optical rotator. Due to the ORD effect, the polarization planes of different spectral components become angularly dispersed after passage through the rotator. The emerging light is partially blocked by the output polarizer and the attenuation of the light at different wavelengths is determined by the material-dependent ORD function. Each wavelength component contributes to every point in the image according to the cosine square of the angle between the polarization of the rotated wavelength component and the fixed output polarization analyzer.

Fig. 5. Absorption and fluorescence spectra of Rhodamine 6G solution: (1) LES fluorescence, (2) long-wavelength absorption spectrum, (3) short-wavelength absorption spectrum, (4) HES fluorescence. (5) absorption of excited molecules. Upper right: scheme of molecular states and optical processes.

Fig. 6. Absorption spectrum (3); and HES fluorescence spectra of benzanthracene solution at stepwise $(\omega_1 + \omega_2)$ excitation with (1) $\omega_1 = 18,800$ cm^{-1} and (2) $\omega_2 = 14,400$ cm^{-1}.

DETAILED DESCIPTION OF THE INVENTION

Current fluorescence techniques, despite their relatively high sensitivity, are restricted by

fundamental photo-physical processes. For certain fluorophores, fluorescence might not

be sufficiently sensitive to be used for successful identification of single-particle samples.

For example, the typical fluorescence spectra of bacteria do not always provide

sufficiently selective signature of pathogens (R.G. Pinnick, *et al.*, "Real-time

measurement of fluorescence spectra from single airborne biological particles", *Field*

Anat. Chem. Technol. 3, 221 (1999); Scully et al., " FAST CARS: Engineering a laser

spectroscopic technique for a rapid identification of bacterial spores". PNAS, 99, 10994,

(2002)).

The invention provides a novel methodology that overcomes limitations of the

conventional fluorescence sensoring. To increase the fluorescence intensity, we will

employ the effect of enhanced fluorophore absorption/emission rates by *surface plasmon*

resonance (SPR) of nearby metal (silver, gold) nanoparticles (M. Kerker, "Optics of

colloid silver", *J. Colloid Interface Sci.* 105, 298 (1985); Lakowicz et al, "Intrinsic

fluorescence from DNA can be enhanced by metallic particles", *Biochem. Biophys. Res.*

Comm. 286, 875 (2001); Gryczynski et al., "Multiphoton excitation of fluorescence near

metallic particles: enhanced and localized excitation", *J. Phys. Chem. B*, 106, 2191

(2002)). When the fluorophore is in a direct contact with a metal nanoparticle,

fluorescence is completely quenched by energy transfer to metal. However, at the

distance of 10 nm – 100s nm between the fluorophore and metal nanoparticle the

absorption and emission rates can be, respectively, enhanced by factors of $\sim 10^2$ and $\sim 10^3$

[11]. The enhancement of the emission intensity depends on fluorescence quantum yield

Q, where $0 \leq Q \leq 1$.

It is the first invention that implements a measurement of multi-band fluorescence for

analyte identification in fluorescence sensing. Current fluorescence sensors are based on

a fundamental principle of molecular fluorescence known as Kasha rule (M. Kasha,

"Characterization of electronic transitions in complex molecules", *Discuss. Faraday Soc.*,

8, 14 (1950)). According to the Kasha rule, a fluorophore in the condensed phase emits a

single-band spectrum from its lowest singlet excited state (LES), due to the vibrational

relaxation and non-radiative dissipation of excitation energy. Natural emission rate for a

fluorophore ($< 10^9$ s^{-1}) defined by fluorophore transient dipole puts a limit on a rate for

fluorophore nonradiative decay of measured fluorescence.

Fluorescence from high-excited state (HES) can provide additional to LES fluorescence

information about molecular structure of analyte in question. However, the non-radiative

decay of the high-excited state is thousands of times faster than HES radiative decay,

which leads to a very low Q for the HES emission (much lower than for the LES

emission) and difficulties in detection of HES fluorescence. The ratio of Q for LES to

HES fluorescence may be as high as 10^5 (Bogdanov, "Fluorescence and multiwave

mixing induced by photon absorption of excited molecules", *Topics in Fluorescence*

Spectroscopy, Vol. 5: *Nonlinear and Two-photon induced Fluorescence*, Ed. J. Lakowicz,

Plenum Press, 1997; Galanin et al. "Fluorescence from the second excited electronic level

and absorption by excited R6G molecules", *Bull. Acad. Sc., Phys. Ser.* 36, 850 (1972)).

Although HES fluorescence is not available in current fluorescence sensing, its

characteristics have sensitivity to both the excitation energy and the fluorophore's

chemical environment. The normally low value of Q prevents the multi-band HES

fluorescence from being used *as a very selective optical signature*. Fig 1 shows

fluorophore electronic states and origin of LES and HES fluorescence spectra.

It is proposed by the invention that measurement of multi-band, LES and HES

fluorescence enhanced by nearby metal nanoparticles can be used as a novel method to

detect an optical signature of sensor and/or analyte. The proposal is based on a discussed

above low Q values for non-enhanced HES fluorescence and observed dependence of

fluorescence enhance effect on fluorophore quantum yield. In a recent experiment, the

emission intensity measured for a series of fluorophores in the vicinity of metal

nanoparticles was greatly increasing for the decreasing values of Q for fluorescence

(Fig.2). This result is consistent with findings observed by Lakowicz *et al.*, noting a

substantial intrinsic fluorescence enhancement for DNA ($Q \approx 0.01\%$) at room

temperature (J.R. Lakowicz et al, "Intrinsic fluorescence from DNA can be enhanced by

metallic particles", *Biochem. Biophys. Res. Comm.* 286, 875 (2001)). Without SPR-

mediated enhancement, the DNA fluorescence could not be observed at room

temperature. Thus, SPR-mediated fluorescence enhancement is a quantum yield

dependent effect.

Because quantum yield and lifetime for HES and LES fluorescence of the same fluorophore differ by orders of magnitude, the enhancement effect is expected to be high for a short-living HES (low Q) and low for a long-living LES (high Q) of the same fluorophore. As a result, fluorescence intensities from HES and LES would reach comparable levels. HES fluorescence could then be used as an additional, new measurable optical signature.

In addition to a better specificity, the sensor proposed in this invention has superior conventional sensors in sensitivity. It is a result of enhance fluorophore absorption rate with nearby metal nanoparticles. Absorption rate enhancement is caused by the electro-magnetic (EM) field E generated by surface plasmons in the evanescent zone. A magnitude of SPR EM field exceeds a magnitude EM field of incident light 10^2 folds. Since the rate of the one-photon excitation is proportional to $|E|^2$, absorption rate can be enhanced by $\sim 10^4$ compare to sensors that do not employ SPR.

Enhanced EM field by surface plasmon is especially effective in non-linear, multi-photon excitation. For a two-photon excitation the absorption rate enhancement could be as high as 10^8 (J. R. Lakowicz, Y. Shen, S. D'Auria, J. Malicka, J. Fang, Z. Gryczynski, and I. Gryczynksi , "Radiative Decay Engineering. 2. Effects of Silver Island Films on Fluorescence Intensity, Lifetimes, and Resonance Energy Transfer", Anal. Biochem., 301:261 (2002)). Metal nanoparticles can also enhance the rate of transient absorption by excited fluorophores in a resonant two-photon HES excitation (M.D.Galanin, and Z.A.

Chizhikova, "Fluorescence from the second excited electronic level and absorption by

excited R6G molecules", *Bull. Acad. Sc., Phys. Ser.* 36, 850 (1972)). The first photon

excites the long-living LES and then, the second photon populates the HES through the

SPR-enhanced absorption. Such a step-wise two-photon HES excitation (M.D.Galanin,

and Z.A. Chizhikova, "Fluorescence from the second excited electronic level and

absorption by excited R6G molecules", *Bull. Acad. Sc., Phys. Ser.* 36, 850 (1972)) is then

followed by the SPR-enhanced emission. The resonance- and SPR-enhanced two-photon

excitation will greatly increase the intensity of the SPR-enhanced HES emission. This is

our concept behind the proposed multi-signature (HES + LES bands) fluorescence

sensing.

Two-photon, step-wise HES excitation has been shown to generate a measurable intensity

of HES emission and to reduce the background contribution (M.D.Galanin and Z.A.

Chizhikova, "Fluorescence from the second excited electronic level and absorption by

excited R6G molecules", *Bull. Acad. Sc., Phys. Ser.* 36, 850 (1972); Lin, and M.R. Topp,

"Low quantum-yield molecular fluorescence: excitation energy dependence and

fluorescence polarization in xanthene dyes", *Chem. Phys. Lett.* 47, 442 (1977)). An

example of HES emission spectrum measured at step-wise excitation of Rhodamine 6G

(R6G) solution is shown in Fig. 5. The position of the short-wavelength HES

fluorescence band correlates with the position of the absorption band but there is no strict

mirror symmetry for these bands. This lack of symmetry is caused by the short HES

fluorescence lifetime (0.2 ps for R6G), as the HES decay competes with vibrational

relaxation of excited fluorophore (M.D.Galanin and Z.A. Chizhikova, "Fluorescence

from the second excited electronic level and absorption by excited R6G molecules", *Bull.*

Acad. Sc., Phys. Ser. 36, 850 (1972).

Another surprising feature of HES emission is the dependence of the fluorescence

spectrum on energy of the excitation photon. Figure 6 shows this dependence for HES

fluorescence of 1,2-benzanthracene solution (Bogdanov, "Fluorescence and multiwave

mixing induced by photon absorption of excited molecules", *Topics in Fluorescence*

Spectroscopy, Vol. 5: *Nonlinear and Two-photon induced Fluorescence*, Ed. J. Lakowicz,

Plenum Press, 1997). Lowering the excitation energy results in both blue shift and more

structure in the HES emission spectrum (the long wavelength LES fluorescence spectrum

is independent of the excitation energy, due to a relatively long LES lifetime). These

results indicate that the HES fluorescence spectrum should be also more sensitive to the

fluorophore's environment and thus provide a more selective spectral signature than the

LES spectrum.

This invention also applied to dual sensing of analytes by using surface enhanced

multiband fluorescence and surface enhanced multiband Raman scattering.

It will be understood by those skilled in the art that the present invention is a novel and

useful method for highly specific, sensitive and fast optochemical sensing.